



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/675,932	09/30/2003	Bruce A. Tockman	GUID.055PA (02-015)	2377

51294 7590 04/11/2006

HOLLINGSWORTH & FUNK, LLC
8009 34TH AVE S.
SUITE 125
MINNEAPOLIS, MN 55425

EXAMINER

SMITH, STEPHANIE R

ART UNIT	PAPER NUMBER
----------	--------------

3762

DATE MAILED: 04/11/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary	Application No. 10/675,932	Applicant(s) TOCKMAN ET AL.	
	Examiner Stephanie Smith	Art Unit 3762	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 1 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 30 September 2003.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-84 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☐ Claim(s) _____ is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☒ Claim(s) 1-84 are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- * See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

Election/Restrictions

Restriction to one of the following inventions is required under 35 U.S.C. 121:

- I. Claims 1-44, drawn to a cardiac lead system that has a first and second stop feature, classified in class 607, subclass 122.
- II. Claims 45-64, drawn to a cardiac lead system having a stop mechanism, classified in class 607, subclass 119.
- III. Claims 65-80, drawn to a method for advancing a cardiac lead system, classified in class 607, subclass 116.
- IV. Claims 81-84, drawn to a means for advancing a cardiac lead system not comprising a lead, classified in class 607, subclass 37.

Inventions II and I are related as combination and subcombination. Inventions in this relationship are distinct if it can be shown that (1) the combination as claimed does not require the particulars of the subcombination as claimed for patentability, and (2) that the subcombination has utility by itself or in other combinations (MPEP § 806.05(c)). In the instant case, the combination as claimed does not require the particulars of the subcombination as claimed because the combination does not require a first stop feature, or portion, and a second stop feature, or portion, wherein engagement of the first stop feature, or portion, and the second stop feature, or portion, prevents further advancement of the elongated body through the cardiac lead lumen. The combination as claimed additionally does not require a first portion and a second portion, the second portion of the cardiac lead lumen distal the first portion and giving

Art Unit: 3762

an inner diameter smaller than the first portion, and wherein engagement of the distal end of the elongated body and the transition region of the cardiac lead lumen prevents further advancement of the elongated body through the cardiac lead lumen. The subcombination has separate utility such as not requiring an inflation member that prevents further advancement of the guide member through the cardiac lead lumen but by engaging the distal end of the elongated body and the transition region of the cardiac lead lumen and also not requiring a separate activating stop mechanism, but two separate mechanisms, one on the lead, one on the member.

Inventions I and II (product) and III are related as product and process of use. The inventions can be shown to be distinct if either or both of the following can be shown: (1) the process for using the product as claimed can be practiced with another materially different product or (2) the product as claimed can be used in a materially different process of using that product. See MPEP § 806.05(h). In the instant case, the product as claimed can be used in a materially different process of using that product, not requiring advancing the lead and member together but separately or using the stop mechanism as a push point, but using the mechanism to prevent puncturing of the lead.

Inventions I and II (combination) and IV are related as combination and subcombination. Inventions in this relationship are distinct if it can be shown that (1) the combination as claimed does not require the particulars of the subcombination as claimed for patentability, and (2) that the subcombination has utility by itself or in other combinations (MPEP § 806.05(c)). In the instant case, the combination as claimed does not require the particulars of the subcombination as claimed because the

Art Unit: 3762

combination as claimed does not require a means for moving the cardiac lead and guide member into a destination vessel or to provide a push point. The subcombination has separate utility such as not requiring a cardiac lead with a conductor but for use with a spinal lead or a cardiac lead without a conductor but using fiber optics.

Inventions III and IV are related as product and process of use. The inventions can be shown to be distinct if either or both of the following can be shown: (1) the process for using the product as claimed can be practiced with another materially different product or (2) the product as claimed can be used in a materially different process of using that product. See MPEP § 806.05(h). In the instant case, the process for using the product as claimed can be practiced with another materially different product not requiring a cardiac lead with a conductor for use with a spinal lead or a cardiac lead without a conductor but using fiber optics.

Because these inventions are independent or distinct for the reasons given above and the inventions require a different field of search (see MPEP § 808.02), restriction for examination purposes as indicated is proper.

A telephone call was made to Mark Hollingsworth on April 10, 2006 to request an oral election to the above restriction requirement, but did not result in an election being made.

Applicant is advised that the reply to this requirement to be complete must include (i) an election of a species or invention to be examined even though the

Art Unit: 3762

requirement be traversed (37 CFR 1.143) and (ii) identification of the claims encompassing the elected invention.

The election of an invention or species may be made with or without traverse. To reserve a right to petition, the election must be made with traverse. If the reply does not distinctly and specifically point out supposed errors in the restriction requirement, the election shall be treated as an election without traverse.

Should applicant traverse on the ground that the inventions or species are not patentably distinct, applicant should submit evidence or identify such evidence now of record showing the inventions or species to be obvious variants or clearly admit on the record that this is the case. In either instance, if the examiner finds one of the inventions unpatentable over the prior art, the evidence or admission may be used in a rejection under 35 U.S.C.103(a) of the other invention.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Stephanie Smith whose telephone number is 571-272-2834. The examiner can normally be reached on Monday-Friday between 7:30 am-4:00 pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Angela Sykes can be reached on 571-272-4955. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3762

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

SRS 4/10/2006
SRS

GEORGE R. EVANISKO
PRIMARY EXAMINER
4/10/6